# R&S®EPL1000 EMI TEST RECEIVER



Specifications



# ROHDE&SCHWARZ

Make ideas real



### CONTENTS

Definitions	3
Specifications	4
Frequency	
Preselection and preamplifier	5
IF and resolution bandwidths	5
Level	5
Sensitivity	6
Measurement speed	
Trigger functions	
I/Q data	9
Audio demodulation	9
Inputs and outputs	9
General data	10
Options	12
R&S <sup>®</sup> FPL1-B30 DC power input 12 V/24 V	
R&S <sup>®</sup> FPL1-B31 internal lithium-ion battery	
R&S <sup>®</sup> FSV-B34 charger (only needed for charging spare batteries)	
R&S <sup>®</sup> EPL1-B91 internal generator	
R&S <sup>®</sup> EPL1-K56 IF analysis	
Ordering information	15
Options	15
Recommended extras	


# Definitions

#### General

Product data applies under the following conditions:

- · Three hours storage at ambient temperature followed by 30 minutes warm-up operation
- Specified environmental conditions met
- · Recommended calibration interval adhered to
- · All internal automatic adjustments performed, if applicable

#### Specifications with limits



#### Non-traceable specifications with limits (n. trc.)

Represent product performance that is specified and tested as described under "Specifications with limits" above. However, product performance in this case cannot be warranted due to the lack of measuring equipment traceable to national metrology standards. In this case, measurements are referenced to standards used in the Rohde & Schwarz laboratories.

#### Specifications without limits

Represent warranted product performance for the specified parameter. These specifications are not specially marked and represent values with no or negligible deviations from the given value (e.g. dimensions or resolution of a setting parameter). Compliance is ensured by design.

#### Typical data (typ.)

Characterizes product performance by means of representative information for the given parameter. When marked with <, > or as a range, it represents the performance met by approximately 80 % of the instruments at production time. Otherwise, it represents the mean value.

#### Nominal values (nom.)

Characterize product performance by means of a representative value for the given parameter (e.g. nominal impedance). In contrast to typical data, a statistical evaluation does not take place and the parameter is not tested during production.

#### Measured values (meas.)

Characterize expected product performance by means of measurement results gained from individual samples.

#### Uncertainties

Represent limits of measurement uncertainty for a given measurand. Uncertainty is defined with a coverage factor of 2 and has been calculated in line with the rules of the Guide to the Expression of Uncertainty in Measurement (GUM), taking into account environmental conditions, aging, wear and tear.

Device settings and GUI parameters are designated with the format "parameter: value".

Non-traceable specifications with limits, typical data as well as nominal and measured values are not warranted by Rohde & Schwarz.

In line with the 3GPP standard, chip rates are specified in million chips per second (Mcps), whereas bit rates and symbol rates are specified in billion bit per second (Gbps), million bit per second (Mbps), thousand bit per second (kbps), million symbols per second (Msps) or thousand symbols per second (ksps), and sample rates are specified in million samples per second (Msample/s). Gbps, Mcps, Mbps, Msps, kbps, ksps and Msample/s are not SI units.

# **Specifications**

Operating modes	receiver mode
	analyzer mode

### Frequency

Frequency range		5 kHz to 30 MHz
Frequency resolution		0.01 Hz
Scaling		linear, logarithmic 1
Reference frequency, internal, nominal	I	
Accuracy		(time since last adjustment x aging rate)
		+ temperature drift + calibration accuracy
Aging per year	standard	1 × 10 <sup>-6</sup>
	with R&S <sup>®</sup> FPL1-B4 OCXO reference	1 × 10 <sup>-7</sup>
	frequency option	
Temperature drift (0 °C to +50 °C)	standard	1 × 10 <sup>-6</sup>
	with R&S <sup>®</sup> FPL1-B4 OCXO reference	1 × 10 <sup>-7</sup>
	frequency option	
Achievable initial calibration accuracy	standard	5 × 10 <sup>-7</sup>
	with R&S <sup>®</sup> FPL1-B4 OCXO reference	5 × 10 <sup>-8</sup>
	frequency option	
Receiver scan		
Scan		max. 10 subranges with different settings
Scan types		stepped, time domain
Measurement time	stepped scan, per frequency	50 µs to 100 s
	time domain scan, per subrange	50 µs to 100 s
Number of trace points		10 000 000
Frequency step size	stepped scan	min. 1 Hz
	time domain scan	$0.25 \times \text{resolution bandwidth}$
Time domain scan		
Frequency segment processed in parallel	RBW = 200 Hz	0.66 MHz
	RBW = 9 kHz	29.85 MHz
	RBW = 120 kHz	24.6 MHz
	RBW = 1 MHz	25.6 MHz
FFT overlap factor		≥ 93 %
Frequency readout (analyzer mode)	l	
Marker resolution		0.01 Hz
Uncertainty		±(marker frequency × reference
		uncertainty + 10 % × resolution bandwidth
		+ $\frac{1}{2}$ (span/(sweep points - 1)) + 1 Hz)
Number of sweep (trace) points	default value	1001
	range	101 to 100001
Marker tuning frequency step size	marker step size = sweep points	span / (sweep points – 1)
	marker step size = standard	span / (default sweep points - 1)
Frequency counter resolution		1 Hz
Count accuracy		±(frequency × reference uncertainty +
, , , , , , , , , , , , , , , , , , ,		1/2 (last digit))
Display range for frequency axis		0 Hz, 10 Hz to maximum frequency
Resolution		0.1 Hz
Maximum span deviation		0.1 %
Sweep time range	span = 0 Hz	1 µs to 8000 s
	span $\ge$ 10 Hz, RBW $\ge$ 100 kHz	1 ms to 8000 s <sup>2</sup>
	span ≥ 10 Hz, RBW < 100 kHz	75 µs to 8000 s <sup>3</sup>
Sweep time accuracy	span = $0 \text{ Hz}$	0.1 % (nom.)
	span ≥ 10 Hz, RBW ≥ 100 kHz	3 % (nom.)

<sup>&</sup>lt;sup>1</sup> Not with internal generator in tracking mode.

 $<sup>^{2}</sup>$   $\,$  Net sweep time without additional hardware settling time.

<sup>&</sup>lt;sup>3</sup> Time for data acquisition for FFT calculation.

### Preselection and preamplifier

Preselection		
State	receiver mode	always on
	analyzer mode	on/off (selectable)
Number of preselection filters		2
Bandwidths (–6 dB), nominal	10 Hz to 150 kHz	fixed lowpass filter
	150 kHz to 30 MHz	38 MHz, fixed bandpass filter
Preamplifier	switchable	
Location		in the signal path between preselection
		and first mixer
Gain		20 dB (nom.)

### IF and resolution bandwidths

EMI filters		
Bandwidths (–6 dB)		10/100/200 Hz, 1/9/10/100/120 kHz, 1 MHz
Bandwidth uncertainty		< 3 %
Shape factor 60 dB:6 dB		< 4
Sweep filters and FFT filters		
Resolution bandwidths (-3 dB)	sweep filters	100 kHz to 10 MHz in 1/2/3/5 sequence
	FFT filters	1 Hz to 50 kHz in 1/2/3/5 sequence
Bandwidth uncertainty		< 3 % (nom.)
Shape factor 60 dB:3 dB		< 5 (nom.)
Channel filters (analyzer mode)		
Bandwidths (–3 dB)		100/200/300/500 Hz, 1/1.5/2/2.4/2.7/3/3.4/4/4.5/5/6/7.5/8.5/9/ 10/12.5/14/15/16/20/21/25/30/50/100/ 150/192/200/300/500 kHz, 1/1.228/1.5/2/3/3.75/5/10 MHz
Bandwidth uncertainty		< 2 % (nom.)
Shape factor 60 dB:3 dB		< 2 (nom.)
Video bandwidths	analyzer mode	1 Hz to 10 MHz in 1/2/3/5 sequence
Signal analysis bandwidth (equalized)	standard, analyzer mode	10 MHz (nom.)

### Level

Display range		displayed noise floor up to +30 dBm	
Maximum input level			
DC voltage	input	0 V	
CW RF power	RF attenuation = 0 dB		
	RF preamplifier off	20 dBm (= 0.1 W)	
	RF preamplifier on	13 dBm (= 0.02 W)	
	RF attenuation ≥ 10 dB		
	RF preamplifier off	30 dBm (= 1 W)	
	RF preamplifier on	23 dBm (= 0.2 W)	
Pulse spectral density	RF attenuation = 0 dB, preselection on,	97 dB μV/MHz	
	RF preamplifier off		
Maximum pulse voltage	RF attenuation ≥ 10 dB	RF attenuation ≥ 10 dB	
	input	450 V	
Maximum pulse energy	RF attenuation ≥ 10 dB, 10 µs	RF attenuation ≥ 10 dB, 10 µs	
	input	20 mWs	
Intermodulation			
1 dB compression	RF attenuation = 0 dB, preselection off,	+10 dBm (nom.)	
(two-tone)	RF preamplifier off		
	RF attenuation = 0 dB, preselection on,	+ 10 dBm (nom.)	
	RF preamplifier off		
	RF attenuation = 0 dB, preselection on,	–15 dBm (nom.)	
	RF preamplifier on		

Third-order intercept point (TOI)	RF attenuation = 0 dB, preselection on, RF preamplifier off,	
	level = $2 \times -20$ dBm, $\Delta f > 5 \times RBW$ or 10 kl	Hz, whichever is larger
	f <sub>in</sub> < 10 MHz	20 dBm (nom.)
	$10 \text{ MHz} \le f_{in} < 30 \text{ MHz}$	> 15 dBm, 20 dBm (typ.)
	RF attenuation = 0 dB, preselection off <sup>4</sup> , R	F preamplifier off,
	level = $2 \times -20$ dBm, $\Delta f > 5 \times$ RBW or 10 kHz, whichever is larger	
	f <sub>in</sub> < 10 MHz	20 dBm (nom.)
	$10 \text{ MHz} \le f_{in} < 30 \text{ MHz}$	> 20 dBm, 23 dBm (typ.)
	RF attenuation = 0 dB, preselection on, RF	preamplifier on,
	level = 2 x –45 dBm, $\Delta f$ > 5 x RBW or 10 kHz, whichever is larger	
	f <sub>in</sub> < 10 MHz	-10 dBm (nom.)
	10 MHz ≤ f <sub>in</sub> < 30 MHz	> –8 dBm
Second-harmonic intercept (SHI)	RF attenuation = 0 dB, level = $-13$ dBm, preselection off, RF preamplifier off	
	$1 \text{ MHz} < f_{in} \le 30 \text{ MHz}$	45 dBm (nom.)

# Sensitivity

Noise indication (receiver mode)	RF attenuation = 0 dB, RF preamplifier off,	termination = 50 $\Omega$ , average detector (AV
	9 kHz ≤ f < 100 kHz, BW = 200 Hz	< –15 dBµV
	100 kHz ≤ f < 150 kHz, BW = 200 Hz	< −15 dBµV
	150 kHz ≤ f < 1 MHz, BW = 9 kHz	< +1 dBµV
	1 MHz ≤ f < 30 MHz, BW = 9 kHz	<-4 dBµV
	RF attenuation = 0 dB, RF preamplifier on,	termination = 50 $\Omega$ , average detector (AV
	9 kHz ≤ f < 100 kHz, BW = 200 Hz	< –25 dBµV
	100 kHz ≤ f < 150 kHz, BW = 200 Hz	< –25 dBµV
	150 kHz ≤ f < 1 MHz, BW = 9 kHz	< –9 dBµV
	1 MHz ≤ f < 10 MHz, BW = 9 kHz	< –16 dBµV
	10 MHz ≤ f < 30 MHz, BW = 9 kHz	< –12 dBµV
Displayed average noise level	RF attenuation = 0 dB, termination = 50 $\Omega$	logarithmic scaling,
DANL, analyzer mode)	normalized to 1 Hz RBW, RBW = 1 kHz, V	
	+20 °C to +30 °C	
	RF preamplifier off, preselection off	
	5 kHz ≤ f < 1 MHz	< -145 dBm, -152 dBm (typ.)
	1 MHz ≤ f < 30 MHz	< -150 dBm, -155 dBm (typ.)
	RF preamplifier off, preselection on	
	5 kHz ≤ f < 1 MHz	< -142 dBm, -147 dBm (typ.)
	1 MHz ≤ f < 30 MHz	< -142 dBm, -147 dBm (typ.)
	RF preamplifier on, preselection on (gain: 20 dB (nom.))	
	1 MHz ≤ f < 10 MHz	< -155 dBm, -158 dBm (typ.)
	10 MHz ≤ f < 30 MHz	< -152 dBm, -156 dBm (typ.)
Spurious responses	input level ≤ –13 dBm, sweep optimization	
Residual spurious response	RF attenuation = 0 dB	, , ,
	f ≤ 2 MHz	< -90 dBm (nom.)
	2 MHz ≤ f < 30 MHz	< –110 dBm
_evel display (receiver mode)		
_evel display	analog	bargraph display, separately for each detector
	digital	numeric, 0.01 dB resolution
Detectors	maximum 4 selectable	maximum peak, minimum peak, RMS,
		average, quasi-peak, CISPR-average,
		RMS-average
Jnits of level axis		dBm, dBµV, dBmV, dBµA, dBpW, dBpT
RF spectrum		
		10 dB to 200 dB, in steps of 10
Logarithmic level axis		
Logarithmic level axis Frequency axis		linear or logarithmic
		linear or logarithmic 6
Frequency axis		6
Frequency axis Number of traces		3

<sup>&</sup>lt;sup>4</sup> Preselection off is only available in analyzer mode. In receiver mode the preselection is permanently on.

Level display (analyzer mode)		
Logarithmic level axis		1 dB to 200 dB, in 1 dB steps
Linear level axis		10 % of reference level per level division, 10 divisions or logarithmic scaling
Number of traces		6
Trace detector		maximum peak, minimum peak, auto peal (normal), sample, RMS, average
Trace functions		clear/write, maximum hold, minimum hold, average, view
EMI detectors		quasi-peak, RMS-average, CISPR average
Measurement marker detector		maximum peak, average, quasi-peak, RMS-average, CISPR-average
Setting range of reference level		<ul> <li>-130 dBm to (-13 dBm + RF attenuation</li> <li>- RF preamplifier gain),</li> <li>in steps of 0.01 dB</li> </ul>
Units of level axis		dBm, dBµV, dBmV, dBµA, dBpW, V, A, W
Level measurement uncertainty		
Absolute level uncertainty at 16.667 MHz	RBW = 10 kHz, level = -10 dBm, reference	e level = $-10 \text{ dBm}$ , RF attenuation = $10 \text{ dB}$
-	+20 °C to +30 °C	
	preselection off	< 0.3 dB (σ = 0.1 dB)
	preselection on	< 0.4 dB (σ = 0.14 dB)
	0 °C to +50 °C	
	preselection off	< 0.5 dB (σ = 0.17 dB)
	preselection on	$< 0.6 \text{ dB} (\sigma = 0.2 \text{ dB})$
Frequency response	RF attenuation = 10/20/30/40 dB, preselec	tion off, RF preamplifier off,
referenced to 16.667 MHz	+20 °C to +30 °C	
	5 kHz ≤ f < 9 kHz	< 1 dB (nom.)
	9 kHz ≤ f < 30 MHz	< 0.3 dB (σ = 0.1 dB)
	RF attenuation = 10/20/30/40 dB, preselec	tion on, RF preamplifier off,
	+20 °C to +30 °C	
	5 kHz ≤ f < 9 kHz	< 1 dB (nom.)
	9 kHz ≤ f < 30 MHz	< 0.8 dB (σ = 0.27 dB)
	any setting of RF attenuation and preselect	tion, RF preamplifier off, 0 °C to +50 °C
	5 kHz ≤ f < 30 MHz	< 1 dB (nom.)
	RF attenuation ≤ 20 dB, RF preamplifier or	
	5 kHz ≤ f < 9 kHz	< 1 dB (nom.)
	9 kHz ≤ f < 30 MHz	$< 0.8 \text{ dB} (\sigma = 0.27 \text{ dB})$
Attenuator switching uncertainty	f = 50 MHz, 0 dB to 55 dB,	< 0.2 dB (σ = 0.07 dB)
	referenced to 10 dB attenuation	
Uncertainty of reference level setting		0 dB <sup>5</sup>
Bandwidth switching uncertainty	referenced to RBW = 10 kHz and sweep ty	
	sweep type = FFT (RBW < 100 kHz)	< 0.1 dB (nom.)
	sweep type = sweep (RBW ≥ 100 kHz)	< 0.2 dB (nom.)
Nonlinearity of displayed level		1
Logarithmic level display	S/N > 16 dB, 0 dB to –50 dB	< 0.1 dB (σ = 0.07 dB)
Linear level display	S/N > 16 dB, 0 dB to -70 dB	5 % of reference level (nom.)
CISPR detectors	CISPR band A/B	in line with CISPR 16-1-1:2019
Total measurement uncertainty	signal level from 0 dB to -50 dB below refe	
	sweep time = auto, sweep type = FFT, RF attenuation = 10/20/30/40 dB, preselection off, RF preamplifier off, span/RBW < 100, confidence level = 95 %, +20 °C to +30 °C	
	1 MHz ≤ f < 30 MHz	0.5 dB
	signal level from 0 dB to -50 dB below refe	erence level, S/N > 20 dB,
	Signal level norm of ab to foo ab below rele	
	sweep time = auto, sweep type = FFT, RF preselection on, RF preamplifier off/on, spa +20 °C to +30 °C	

<sup>&</sup>lt;sup>5</sup> The setting of the reference level affects only the graphical representation of the measurement result on the display, not the measurement itself. Therefore, the reference level setting causes no additional uncertainty in measurement results.

### **Measurement speed**

Receiver mode Time domain scan	CISPR band A, 9 kHz to 150 kHz,	500 ms (meas.)
	RBW = 200  Hz or	
	CISPR band B, 150 kHz to 30 MHz,	
	RBW = 9  kHz:	
	measurement time = $100 \text{ ms}$ ,	
	peak detector	
	CISPR band A, 9 kHz to 150 kHz,	1.4 s (meas.)
	RBW = 200 Hz or	
	CISPR band B, 150 kHz to 30 MHz,	
	RBW = 9 kHz;	
	measurement time = 1 s,	
	peak detector	
	CISPR band A, 9 kHz to 150 kHz,	≤3 s (meas.)
	RBW = 200 Hz or	
	CISPR band B, 150 kHz to 30 MHz,	
	RBW = 9 kHz;	
	measurement time = 1 s,	
	quasipeak and CISPR-average detector	
	CISPR band A, 9 kHz to 150 kHz,	15.4 s (meas.)
	RBW = 200 Hz or	
	CISPR band B, 150 kHz to 30 MHz,	
	RBW = 9 kHz;	
	measurement time = 15 s,	
	peak detector	
	CISPR band A, 9 kHz to 150 kHz,	≤17 s (meas.)
	RBW = 200 Hz or	
	CISPR band B, 150 kHz to 30 MHz,	
	RBW = 9 kHz;	
	measurement time = 15 s,	
Local measurement and display update	quasipeak and CISPR-average detector 1001 sweep points, sweep optimization	5 ms (200/s) (nom.)
rate	set to "speed"	
Maximum sweep rate,	trace average = on	0.9 ms (1100/s) (nom.)
remote operation <sup>6, 7</sup>		$2.2 m_{2} (2EZ/2) (n - m)$
Remote measurement and LAN transfer <sup>6</sup>		3.2 ms (357/s) (nom.)
Marker peak search <sup>6</sup>		1.9 ms (nom.)
Center frequency tune + sweep + sweep data transfer <sup>6</sup>		16 ms (nom.)

# **Trigger functions**

Trigger		
Trigger source		free run, video, external, IF power, I/Q power
Trigger offset	span ≥ 10 Hz	0 s to 20 s
	span = 0 Hz	(-sweep time) to 20 s
Maximum deviation of trigger offset		±10 ns
IF power trigger (analyzer mode)		
Sensitivity	minimum signal power	<ul> <li>–60 dBm + RF attenuation –</li> <li>RF preamplifier gain</li> </ul>
	maximum signal power	–15 dBm + RF attenuation – RF preamplifier gain
IF power trigger bandwidth		10 MHz (nom.)
Gated sweep		
Gate source		video, external, IF power,
Gate delay		0 s to 20 s,
		minimum resolution: 10 ns
Gate length		10 ns to 20 s,
		minimum resolution: 10 ns
Maximum deviation of gate length		±10 ns

 $<sup>^{\</sup>rm 6}$   $\,$  Measured with a PC equipped with Intel  $^{\rm @}$  Core  $^{\rm TM}$  i7 2.8 GHz and Gigabit LAN interface.

<sup>&</sup>lt;sup>7</sup> Measurement is performed with a sweep count of 1000. The indicated speed is the average speed of 1 sweep.

### I/Q data

Interface		GPIB or LAN interface
Memory length		max. 25 Msample I and Q
Word length of I/Q samples		14 bit
Sampling rate	standard	100 Hz to 16 MHz
Maximum signal analysis bandwidth (equalized)	standard	12.8 MHz
Signal analysis bandwidth ≤ 10 MHz		
Amplitude flatness	f <sub>center</sub> ≥ 12 MHz and (1.25 × signal analysis bandwidth)	±0.3 dB (nom.)
Deviation from linear phase	f <sub>center</sub> ≥ 12 MHz and (1.25 × signal analysis bandwidth)	±1° (nom.)
Signal analysis bandwidth ≤ 30 MHz		,
Amplitude flatness	f <sub>center</sub> ≥ 12 MHz and (1.25 × signal analysis bandwidth)	±0.5 dB (nom.)
Deviation from linear phase	f <sub>center</sub> ≥ 12 MHz and (1.25 × signal analysis bandwidth)	±1.5° (nom.)

### Audio demodulation

AF demodulation types	AM and FM
Audio output	loudspeaker and phone jack
Marker stop time in analyzer mode	100 ms to 60 s

# Inputs and outputs

RF input			
Impedance		50 Ω	
Connector		N female	
VSWR	RF attenuation ≥ 10 dB, receive	RF attenuation ≥ 10 dB, receiver mode or analyzer mode with preselection on	
	9 kHz ≤ f < 30 MHz	< 1.2	
	RF attenuation ≤ 10 dB, receive	r mode or analyzer mode with preselection on	
	9 kHz ≤ f < 30 MHz	< 2.0	
	RF attenuation ≥ 10 dB, analyzer mode with preselection off		
	9 kHz ≤ f < 30 MHz	< 1.5 (nom.)	
Setting range of attenuator	input	0 dB to 55 dB, in 1 dB steps	
USB interfaces	front	2 ports, type A plug, version 2.0	
	rear	2 ports, type A plug, version 3.1	
Reference output			
Connector		BNC female	
Impedance		50 Ω	
Output frequency	internal reference	10 MHz	
	external reference	same as reference input signal	
Level		> 0 dBm (nom.)	
Reference input			
Connector		BNC female	
Impedance		50 Ω	
Input frequency range		10 MHz ± 5 ppm	
Required level		> 0 dBm into 50 Ω	
External trigger/gate input			
Connector		BNC female	
Trigger voltage		0.5 V to 3.5 V	
Input impedance		10 kΩ	
IEC/IEEE bus control		interface in line with IEC 625-2	
		(IEEE 488.2)	
Command set		SCPI 1997.0	
Connector		24 pin Amphenol female	
Interface functions		SH1, AH1, T6, L4, SR1, RL1, PP1, DC1, DT1, C0	
LAN interface		10/100/1000BASE-T	
Connector		RJ-45	
External monitor	1		
Connector		DisplayPort rev. 1.3	

User port		
Connector		25 pin D-Sub female
Output	TTL compatible, 0 V/5 V, max. 15 mA	
Input		TTL compatible, max. 5 V
Noise source control and power sens	or	
Connectors	for R&S <sup>®</sup> NRP-Zxx power sensors	7 pin LEMOSA female
	for noise source control	BNC female
Noise source control output voltage		0 V/28 V, switchable,
		max. 100 mA (nom.)
IF/video/demod out (analyzer mode)		
Connector		BNC female, 50 Ω
IF out		
Bandwidth		equal to RBW setting
IF frequency		25 MHz
Output level	center frequency > 10 MHz, span = 0 Hz,	0 dBm (nom.)
	signal at reference level and center	
	frequency	
Video out		
Bandwidth		equal to VBW setting
Output scaling	logarithmic display scale	logarithmic
	linear display scale	linear
Output level	center frequency > 10 MHz, span = 0 Hz,	1 V (nom.), open circuit
	signal at reference level and center	
	frequency	
Audio output		
Loudspeaker		built-in, volume adjustable
AFout		·
Connector		3.5 mm mini jack
Output impedance		10 Ω
Open-circuit voltage		up to 1.5 V, adjustable

### **General data**

Display		21 cm LC TFT color display (10.1")
Resolution		1280 × 800 pixel (WXGA resolution)
Pixel failure rate		< 1 × 10 <sup>-5</sup>
Data storage		
Internal	standard	solid-state drive (SSD) 128 Gbyte
External		supports USB 2.0/3.1 compatible memory
		devices
Environmental conditions		
Temperature	operating temperature range	+0 °C to +50 °C
	storage temperature range	–20 °C to +70 °C
Climatic loading	without condensation	+40 °C at 85 % relative humidity,
-		in line with EN 60068-2-30,
Mechanical resistance		
Vibration	sinusoidal	5 Hz to 55 Hz,
		0.15 mm constant amplitude
		(1.8 g at 55 Hz);
		55 Hz to 150 Hz,
		acceleration: 0.5 g constant;
		in line with EN 60068-2-6
	random	10 Hz to 300 Hz,
		acceleration: 1.2 g (RMS),
		in line with EN 60068-2-64
Shock		40 g shock spectrum, in line with
		MIL-STD-810E method no. 516.4
		procedure I, MIL-PRF-28800F
EMC		in line with EMC Directive 2014/30/EU
		including IEC/EN 61326-1 <sup>9,10</sup> ,
		IEC/EN 61326-2-1,
		CISPR 11/EN 55011 9,
		IEC/EN 61000-3-2, IEC/EN 61000-3-3

<sup>9</sup> Emission limits for class B equipment.

<sup>10</sup> Immunity test requirement for industrial environment (EN 61326 table 2).

Recommended calibration interval		1 years
Power supply		
AC supply	with battery option	100 V to 240 V ± 10 %,
		50 Hz to 60 Hz ± 5 %
Current consumption	without options	nom. 2.16 A (at 100 V) to
		0.95 A (at 240 V)
	with internal battery	nom. 3 A to 1.5 A
	(R&S <sup>®</sup> FPL1-B31 option in charge mode)	
Safety		in line with EN 61010-1, IEC 61010-1,
		UL 61010-1, CAN/CSA-C22.2
		No. 61010-1
Test mark		CE, KCC, CSA
Dimensions and weight		
Dimensions	W×H×D	408 mm × 186 mm × 235 mm
		(16.06 in × 7.32 in × 9.25 in)
Net weight, nominal	without options	6.9 kg (15.2 lb)
-	with internal battery	8.6 kg (18.95 lb)

# Options

### R&S®FPL1-B30 DC power input 12 V/24 V

Input voltage range	DC	12 V to 24 V (nom.),
		10.4 V to 28 V,
		switch-on voltage > 11 V (meas.)
Input current	V <sub>in</sub> = 12 V/24 V	13 A/6.5 A (nom.)
	$V_{in} = 12 V/24 V$ , operating mode, without	6.8 A/3.2 A (meas.)
	internal batteries (R&S <sup>®</sup> FPL1-B31)	
	$V_{in} = 12 V/24 V$ , operating mode,	11 A/5 A (meas.)
	internal batteries in charge mode	
	$V_{in} = 12 V/24 V$ , instrument standby mode,	6.5 A/3.0 A (meas.)
	internal batteries in charge mode	
Temperature	operating temperature range	0 °C to +40 °C
	storage temperature range	–20 °C to +70 °C

### R&S®FPL1-B31 internal lithium-ion battery

Operating time		2 h (nom.)
Charge time	standby mode, AC supply	< 2 h (nom.)
	standby mode, external DC supply (R&S <sup>®</sup> FPL1-B30)	< 2 h (nom.)
	operating mode	< 4 h (nom.)
Temperature	operating temperature range, discharge	0 °C to +50 °C
	operating temperature range, charge	0 °C to + 45 °C
	storage temperature range	-20 °C to +60 °C <sup>11</sup>

### R&S<sup>®</sup>FSV-B34 charger (only needed for charging spare batteries)


AC input voltage range		100 V to 240 V, ±10 % (nom.)
AC supply frequency		50 Hz to 60 Hz (nom.)
Power consumption		max. 300 W (nom.)
Number of charger bays		4
Dimensions	W×H×D	400 mm × 127 mm × 203 mm
		(15.75 in × 5 in × 8 in)
Net weight		3.1 kg (6.9 lb)

### R&S®EPL1-B91 internal generator


Modes		tracking generator	
		independent source	
		power sweep	
Frequency			
Frequency range		5 kHz to 30 MHz	
Setting resolution	independent CW source	0.01 Hz	
Frequency offset			
Setting range		0 Hz to 30 MHz	
Setting resolution		0.01 Hz	
Spectral purity			
SSB phase noise	frequency = 15 MHz, output level = 0 dBn	n	
	carrier offset = 10 kHz	< -102 dBc (1 Hz), -108 dBc (1 Hz) (typ.)	
	carrier offset = 100 kHz	< –105 dBc (1 Hz), –111 dBc (1 Hz) (typ.)	
	carrier offset = 1 MHz	< –117 dB (1 Hz), –130 dBc (1 Hz) (typ.)	
Harmonics	output level = 0 dBm	output level = 0 dBm	
	5 kHz ≤ f < 100 kHz	< -30 dBc (nom.)	
	100 kHz ≤ f ≤ 30 MHz	< -30 dBc	
Non-harmonic spurious	output level = 0 dBm		
	1 kHz < offset from carrier ≤ 4 MHz	–35 dBc (nom.)	
	offset from carrier > 4 MHz	< -35 dBc, -45 dBc (typ.)	

<sup>&</sup>lt;sup>11</sup> The battery packs should be stored in an environment with low humidity, free from corrosive gas at a recommended temperature range < +21 °C. Extended exposure to temperatures above +45 °C could degrade battery performance and life.

Level		
Specified level range		-50 dBm to 0 dBm
Setting resolution		0.1 dB
Setting range		-60 dBm to +10 dBm
Absolute level uncertainty	frequency =16.667MHz, +20 °C to +30 °C, output level = $-10$ dBm, frequency offset = 0 Hz	< 0.5 dB
Frequency response	output level = -10 dBm, referenced to level at 16.667 MHz, +20 °C to +30 °C,frequency offset = 0 Hz100 kHz $\leq$ f $\leq$ 30 MHz $<$ 1 dB	
Level nonlinearity	for specified level range, referenced to -10 dBm output level, +20 °C to +30 °C, $f \ge 100 \text{ kHz}$	≤ 2 dB, < 0.5 dB (typ.)



Maximum output power versus frequency, level in dBm (meas.)



Minimum output power versus frequency, level in dBm (meas.)

Dynamic range	RBW = 1 kHz, f = 30 MHz	115 dB (nom.)
Power sweep		
Specified level range		-50 dBm to 0 dBm
Setting resolution		0.1 dB
Setting range		-60 dBm to +10 dBm
GEN output		
Connector		N female, 50 Ω
VSWR		1.5 (nom.)
Reverse power		
DC voltage		50 V
CW RF power		30 dBm (= 1 W)
Maximum pulse voltage		150 V
Maximum pulse energy	pulse duration: 10 µs	1 mWs

# R&S<sup>®</sup>EPL1-K56 IF analysis

Level display (receiver mode)	
IF spectrum	
Span	max. 10 MHz
Resolution bandwidths	10 Hz to 100 kHz, in 1/2/3/5 sequence
Detector	sample
Logarithmic level axis	10 dB to 200 dB, in steps of 10 dB
Frequency axis	linear
Number of traces	3

# **Ordering information**

Designation	Туре	Order No.
EMI test receiver	R&S <sup>®</sup> EPL1000	1350.4444.10
Accessories supplied: power cable, quick start quide		

# Options

Designation	Туре	Order No.	Retrofittable	Remarks
OCXO reference frequency	R&S <sup>®</sup> FPL1-B4	1323.1902.02	yes	retrofit in service center
GPIB interface	R&S <sup>®</sup> FPL1-B10	1323.1890.02	yes	user-retrofittable
Replacement SSD including controller unit	R&S <sup>®</sup> EPL1-B19	1350.4450.02	yes	user-retrofittable mounted on PC board, including analyzer firmware
DC power supply, 12 V/24 V	R&S <sup>®</sup> FPL1-B30	1323.1877.02	yes	user-retrofittable
Internal lithium-ion battery	R&S <sup>®</sup> FPL1-B31	1323.1725.02	yes	retrofit in service center; including 2 battery packs and internal charging unit
Internal generator 5 kHz to 30 MHz	R&S <sup>®</sup> EPL1-B91	1350.4073.02	no	
Firmware				
AM/FM/qM measurement demodulator	R&S <sup>®</sup> FPL1-K7	1323.1731.02		
IF analysis	R&S <sup>®</sup> EPL1-K56	1350.4067.02		
EMC test software				
Essential EMI test software	R&S <sup>®</sup> ELEMI-E	5601.0030.02		
License dongle	R&S <sup>®</sup> EMCPC	5601.0018.02		

### **Recommended extras**

Designation	Туре	Order No.
Soft carrying bag for transport and outdoor operation	R&S <sup>®</sup> EPL1-Z2	1350.4309.02
H-style shoulder harness (requires R&S <sup>®</sup> EPL1-Z2 option)	R&S <sup>®</sup> EPL1-Z3	1350.4315.02
Spare lithium-ion battery pack	R&S <sup>®</sup> FPL1-Z4	1323.1677.02
Anti-glare display film for outdoor operation	R&S <sup>®</sup> FPL1-Z5	1323.1690.02
Lithium-ion battery charger for charging spare batteries	R&S <sup>®</sup> FSV-B34	1321.3950.02
19" rackmount kit	R&S®EPL1-Z6	1350.4321.02
Control cable for R&S <sup>®</sup> ENV216/R&S <sup>®</sup> ENV432/R&S <sup>®</sup> ENV420		1
Length: 3 m	R&S®EZ-21	1107.2087.03
Length: 10 m	R&S®EZ-21	1107.2087.10
Accredited calibration	R&S <sup>®</sup> ACAEPL1000	Please contact your local
		Rohde & Schwarz
		sales office.

Service options		
Extended warranty, one year	R&S <sup>®</sup> WE1	Please contact
Extended warranty, two years	R&S <sup>®</sup> WE2	your local
Extended warranty, three years	R&S <sup>®</sup> WE3	Rohde & Schwarz sales office.
Extended warranty, four years	R&S®WE4	
Extended warranty with calibration coverage, one year	R&S <sup>®</sup> CW1	
Extended warranty with calibration coverage, two years	R&S <sup>®</sup> CW2	
Extended warranty with calibration coverage, three years	R&S <sup>®</sup> CW3	
Extended warranty with calibration coverage, four years	R&S <sup>®</sup> CW4	
Extended warranty with accredited calibration coverage, one year	R&S <sup>®</sup> AW1	
Extended warranty with accredited calibration coverage, two years	R&S <sup>®</sup> AW2	
Extended warranty with accredited calibration coverage, three years	R&S®AW3	
Extended warranty with accredited calibration coverage, four years	R&S <sup>®</sup> AW4	

#### Extended warranty with a term of one to four years (WE1 to WE4)

Repairs carried out during the contract term are free of charge <sup>12</sup>. Necessary calibration and adjustments carried out during repairs are also covered.

#### Extended warranty with calibration (CW1 to CW4)

Enhance your extended warranty by adding calibration coverage at a package price. This package ensures that your Rohde & Schwarz product is regularly calibrated, inspected and maintained during the term of the contract. It includes all repairs <sup>12</sup> and calibration at the recommended intervals as well as any calibration carried out during repairs or option upgrades.

#### Extended warranty with accredited calibration (AW1 to AW4)

Enhance your extended warranty by adding accredited calibration coverage at a package price. This package ensures that your Rohde & Schwarz product is regularly calibrated under accreditation, inspected and maintained during the term of the contract. It includes all repairs <sup>12</sup> and accredited calibration at the recommended intervals as well as any accredited calibration carried out during repairs or option upgrades.

<sup>&</sup>lt;sup>12</sup> Excluding defects caused by incorrect operation or handling and force majeure. Wear-and-tear parts are not included.

Version 01.01, March 2023

### Service at Rohde & Schwarz You're in great hands

- ► Worldwide
- Local and personalized
- Customized and flexible
- Uncompromising quality
   Long-term dependability

### Rohde & Schwarz

The Rohde&Schwarz technology group is among the trailblazers when it comes to paving the way for a safer and connected world with its leading solutions in test&measurement, technology systems and networks&cybersecurity. Founded more than 85 years ago, the group is a reliable partner for industry and government customers around the globe. The independent company is headquartered in Munich, Germany and has an extensive sales and service network with locations in more than 70 countries.

www.rohde-schwarz.com

### Sustainable product design

- ► Environmental compatibility and eco-footprint
- Energy efficiency and low emissions
- Longevity and optimized total cost of ownership



#### Rohde & Schwarz training

www.training.rohde-schwarz.com

### Rohde & Schwarz customer support

www.rohde-schwarz.com/support



#### DISTRAME

Parc du Grand Troyes - Quartier Europe Centrale, 40 rue de Vienne - 10300 SAINTE-SAVINE Tél. : 03 25 71 25 83 - infos@distrame.fr - www.distrame.fr

383.7819.22 01.01 PDP/PDW 1 en

R&S° is a registered trademark of Rohde&Schwarz GmbH&Co. KG Trade names are trademarks of the owners PD 3683.7819.22 | Version 01.01 | March 2023 (ch) R&S°EPL1000 EMI Test Receiver Data without tolerance limits is not binding | Subject to change © 2022 - 2023 Rohde&Schwarz GmbH&Co. KG | 81671 Munich, Germany